

Qualitative scoring report

		Roche cobas/Modular/Elecsys Toxo IgG Avidity		4						
		TestLine Toxoplasma IgG Avidity EIA		1						
		Vidia MONO-VIDITEST anti-Toxoplasma IgG avidity		1						
	Total:		50		-	-	-	-		100 %

		Roche cobas/Modular/Elecsys Toxo IgG Avidity		4					
		TestLine Toxoplasma IgG Avidity EIA		1					
		Vidia MONO-VIDITEST anti-Toxoplasma IgG avidity		1					
	Total:		50		-	-	-	-	100 %

Report Info**PARTICIPANTS**

Altogether 155 laboratories from 19 countries participated in this EQA round.

REPORT INFO

The results are divided into groups according to the method stated by the laboratory and presented in laboratory-specific tables. Accepted results are marked with green color and laboratory's own result with a black radio button . In the scoring report you will find summaries of overall success rate and sample specific success rates (%). Sample specific interpretations are shown in pie diagrams as percentages and the total interpretation and methodic counts in the tables. If you have not reported any results you will get a note: "You have not responded in time, only global report is available."

For information on report interpretation and performance evaluation, please see the "EQAS Interpretation guidelines" in LabScala User instructions. In case you have any questions regarding the reports, please contact the EQA Coordinator.

SCORING

The round is scored based on test results and clinical interpretations when 60% or more of the participants report the expected result and when at least three results are reported. The following general rules are applied:

Correct/expected test result 2/2 points
False/deviating test result 0/2 points
Correct/expected clinical interpretation 4/4 points
False/deviating clinical interpretation 0/4 points

The performance of the laboratory is assessed by the Own success rate (%). The target is 100%. The examination-specific scores obtained by the laboratory in the round are converted to Own success rate per sample (scores/maximum scores*100). The Laboratory's Own success rate is the average of the sample success rates. The success rate for the entire round (AVR success rate) is calculated from the total number of scores given to the results per sample (all scores/maximum scores*100). The AVG success rate of the entire round is the average of the sample success rates. The difference in the Laboratory's Own success rate (%) to the corresponding numbers for the entire round is shown in the table.