



## EU Declaration of Conformity for product Anchor Tips, 300uL

REF

10036506

Basic UDI-DI

91201239109889J4



STRATEC Consumables GmbH  
Sonystrasse 20, 5081 Anif, Austria  
SRN: AT-MF-000023689

We, as manufacturer of Anchor Tips, 300uL take sole responsibility for and hereby declare that Anchor Tips, 300uL meets the provisions of the following regulation:

- *REGULATION (EU) 2017/746 on in vitro medical devices*

### RISK Class

A  B  C  D

The Anchor Tips, 300uL are validated for their intended use in combination with the KleeYa System. In their function as KleeYa accessories, they are classified as class A IVD medical devices.

### CONFORMITY ROUTE

- ANNEX IX Technical Documentation Examination  
 ANNEX IX Full Quality System  
 ANNEX XI Production Quality System  
 ANNEX I & II + III

Conformity with the relevant General Safety and Performance Requirements is demonstrated for the Anchor Tips, 300uL in combination with the KleeYa System. The application of the KleeYa System for in vitro diagnostic purposes requires a separate conformity assessment according to REGULATION (EU) 2017/746 for the complete system into which it will be incorporated and / or is used in combination with (e.g. assay).

This Declaration of Conformity is valid for the product's configuration and the regulatory requirements effectual at the date the Declaration was issued. Changes affecting the product, and / or the applicable regulations trigger a review of the conformity assessment the Declaration is based on, and the issuance of a new version of the document.

Date: 2022-05-25

Place: Anif

Signature:

  
Thomas Ehrenfeld  
PRRC  
STRATEC Consumables GmbH

stratec●●  
consumables

STRATEC Consumables GmbH  
5081 Anif, Sonystrasse 20